

CSIR-NIIST/PUR/IMP/009/25

05/02/2026

**Corrigendum w.r.t. Amendment of Tender Specifications & Extension of Bid Submission & Opening Dates**

With reference to the CPPP Tender Ref: 2026\_CSIR\_263147\_1 (CSIR-NIIST File Reference No. CSIR-NIIST/PUR/IMP/009/25), all the prospective bidders are hereby informed that CSIR-NIIST, Thiruvananthapuram has amended the Technical Specifications of Tender and also extended the Bid Submission due date and opening date as detailed below in order to facilitate the prospective bidders to submit the bid against the CPPP tender:

Features/Parameters	Specifications as Existing	Specifications as Amended
Temperature Control	Infrared (IR) sensor or optional fiber-optic probe, with bottom mounted temperature control system, precision of $\pm 1^{\circ}\text{C}$	Infrared (IR) sensor or optional fiber-optic probe, with a temperature control system having precision of $\pm 1^{\circ}\text{C}$
Vessel Compatibility	Minimum 100 mL Pyrex sealed vessels	Minimum 100 mL Pyrex/PTFE/ Polymer/SiC vessels, with a mount. A minimum of 4 vessels of 100 mL or above should be supplied
Operating Power	800 W	800 – 2500 W

Details of Due Date & Opening Date	Existing as :	Amended as :
Bid Submission End date	06/02/2026 03:00 PM	19/02/2026 10.00 AM
Bid Opening date	07/02/2026 03:30 PM	20/02/2026 10.30 AM

All other Specifications and Terms & Conditions will remain the same as per the original tender. Accordingly, the prospective bidders are requested to take notice of the above and submit their bids on or before the bid submission end date as per the tender and corrigendum.

Yours faithfully,

  
5/2/26  
Stores & Purchase Officer  
(For & on behalf of CSIR-INDIA)  
Email: [spo.niist@csir.res.in](mailto:spo.niist@csir.res.in)  
National Institute of Science and Technology  
Tiruvananthapuram/Thiruvananthapuram